

Millipore filter replacement schedule

Dinsmore Lab

Sanitize with Cl: 1/(4 wks.)

Prefilters: replace at least 1/mo (or if the left one turns yellow or there is a pressure difference between the two prefilters.). White one goes on left side. Order replacements if there are none left for the next month.

Cartridges: replace 1/(6 mo) (see alarm on unit). Order replacements immediately after exchanging.

Vent filter: (on top of tank): replace 1/yr. Order replacement as part of next order from Millipore.

[illegible]